
Midmark M11 - v1.0.0
Total Cycles: 4800

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2024 04:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 71.0 F 0.0
4:00 96.2 F 0.1
6:00 126.2 F 0.3
8:00 153.5 F 0.7
10:00 180.9 F 2.8
12:00 196.9 F 5.9
14:00 208.6 F 9.0
16:00 218.3 F 13.1
18:00 224.0 F 7.1
20:00 235.7 F 11.7
22:00 233.8 F 8.4
24:00 241.2 F 11.0
26:00 245.5 F 13.0
28:00 257.7 F 19.5
30:00 268.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.2 F 28.5
2:00 272.7 F 28.9
3:00 272.2 F 28.9
4:00 271.7 F 28.2

Min 271.6 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 168.1 F 0.0
5:00 185.9 F 0.0
10:00 186.0 F 0.0
15:00 181.2 F 0.0
20:00 185.5 F 0.0
25:00 205.9 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 31:05
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:12
CYCLE COMPLETE

08/09/2024 06:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____