
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001861
10:58 16/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:10
STERILIZING 4:28
136.0 C 222kPa 4:28
136.4 C 226kPa 4:58
136.3 C 224kPa 5:28
136.3 C 225kPa 5:58
136.3 C 225kPa 6:28
136.4 C 226kPa 6:58
136.3 C 225kPa 7:28
136.4 C 224kPa 7:58
136.4 C 224kPa 7:58
VENTING 7:58
AIR DRYING 8:56
CYCLE COMPLETE 9:18

Digital Signature #
82F696BC0C63E5D6
