

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2025

Name: RITTER2  
Sterilizer ID: V2313228  
02/08/2023 12:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
2:00 191.4 F 3.6  
4:00 197.7 F 8.2  
6:00 221.5 F 10.5  
8:00 229.8 F 9.1  
10:00 235.6 F 9.6  
12:00 242.8 F 13.1  
14:00 243.6 F 12.3  
16:00 256.5 F 19.2  
18:00 268.4 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.8  
1:00 272.0 F 29.2  
2:00 271.4 F 29.0  
3:00 272.1 F 29.3  
4:00 271.2 F 28.8  
  
Min 271.1 F 28.7  
Max 272.2 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
5:00 230.1 F 0.0  
10:00 233.6 F 0.0  
15:00 229.5 F 0.0  
20:00 228.4 F 0.0  
25:00 231.4 F 0.0  
30:00 227.8 F 0.0  
  
Drying Complete

FILLING: 1:32  
HEATING: 19:09  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:51  
CYCLE COMPLETE

02/08/2023 12:57 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_