

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2384

Name: RITTER2  
Sterilizer ID: V2313228  
05/17/2023 11:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.1 F 0.0  
2:00 185.0 F 3.6  
4:00 199.3 F 8.3  
6:00 220.9 F 10.0  
8:00 230.8 F 9.7  
10:00 234.5 F 8.7  
12:00 244.3 F 13.2  
14:00 244.3 F 12.7  
16:00 257.1 F 19.6  
18:00 268.3 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.2 F 29.9  
2:00 272.8 F 29.9  
3:00 271.6 F 29.3  
4:00 272.4 F 30.1  
  
Min 270.8 F 28.7  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 220.3 F 0.0  
10:00 236.6 F 0.0  
15:00 241.7 F 0.0  
20:00 230.9 F 0.0  
25:00 223.2 F 0.0  
30:00 223.0 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 19:16  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:24  
CYCLE COMPLETE

05/17/2023 11:59 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_