
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 001509
12:37 10/18/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:13
PRESSURIZING 7:08
STERILIZING 8:46
136.0 C 220kPa 8:46
136.2 C 220kPa 9:16
136.4 C 223kPa 9:46
136.2 C 221kPa 10:16
136.2 C 221kPa 10:46
136.3 C 222kPa 11:16
136.3 C 222kPa 11:46
136.5 C 222kPa 12:16
VENTING 12:16
AIR DRYING 13:30
CYCLE COMPLETE 25:55

Digital Signature #
054A68617016993C
