

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2879

Name: RITTER 1  
Sterilizer ID: V2178198  
02/09/2023 09:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 213.5 F 0.0  
2:00 203.8 F 2.9  
4:00 209.8 F 7.3  
6:00 216.4 F 13.0  
8:00 227.6 F 8.0  
10:00 239.3 F 13.4  
12:00 239.4 F 10.9  
14:00 237.3 F 9.2  
16:00 251.5 F 16.1  
18:00 263.3 F 22.9  
20:00 271.8 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.5 F 29.0  
2:00 271.5 F 28.3  
3:00 272.6 F 29.2  
4:00 271.9 F 28.6

Min 271.5 F 28.3  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.9 F 0.0  
5:00 225.8 F 0.0  
10:00 234.3 F 0.0  
15:00 226.5 F 0.0  
20:00 223.9 F 0.0  
25:00 223.2 F 0.0  
30:00 220.1 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 20:08  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:11  
CYCLE COMPLETE

02/09/2023 10:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_