
Midmark M11 - v1.0.5
Total Cycles: 4355

Name: RITTER2
Sterilizer ID: V2313228
02/13/2025 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.1 F 0.0
2:00 110.2 F 0.8
4:00 127.0 F 3.1
6:00 154.8 F 6.5
8:00 180.7 F 10.9
10:00 202.7 F 2.9
12:00 217.6 F 6.8
14:00 230.3 F 11.3
16:00 230.8 F 7.9
18:00 243.1 F 12.7
20:00 243.4 F 12.6
22:00 255.7 F 19.1
24:00 266.8 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.9 F 29.8
2:00 271.8 F 29.3
3:00 272.9 F 29.8
4:00 271.8 F 29.2

Min 271.7 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.8 F 0.0
5:00 183.5 F 0.0
10:00 191.9 F 0.0
15:00 198.2 F 0.0
20:00 203.7 F 0.0
25:00 202.2 F 0.0
30:00 208.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:33
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:03:55
CYCLE COMPLETE

02/13/2025 12:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____