
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005370
13:07 18/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:14
STERILIZING 3:15
136.4 C 223kPa 3:15
136.2 C 224kPa 3:45
136.3 C 224kPa 4:15
136.4 C 223kPa 4:45
136.1 C 219kPa 5:15
136.4 C 224kPa 5:45
136.4 C 224kPa 6:15
136.4 C 224kPa 6:45
136.4 C 224kPa 6:45
VENTING 6:45
AIR DRYING 7:33
CYCLE COMPLETE 7:37

Digital Signature #
A18E317B3229D356
