

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2253

Name: RITTER2  
Sterilizer ID: V2313228  
04/14/2023 10:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.3 F 0.0  
2:00 184.3 F 3.1  
4:00 198.4 F 7.6  
6:00 216.5 F 12.9  
8:00 228.0 F 8.6  
10:00 238.3 F 11.0  
12:00 240.5 F 11.8  
14:00 239.7 F 10.6  
16:00 252.5 F 17.1  
18:00 263.7 F 23.8  
20:00 272.0 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 272.6 F 30.1  
2:00 272.3 F 29.9  
3:00 272.1 F 29.6  
4:00 272.4 F 29.6

Min 271.0 F 28.9  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.0 F 0.0  
5:00 241.0 F 0.0  
10:00 229.7 F 0.0  
15:00 227.7 F 0.0  
20:00 221.4 F 0.0  
25:00 217.5 F 0.0  
30:00 216.7 F 0.0  
Drying Complete

FILLING: 1:59  
HEATING: 20:11  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:21  
CYCLE COMPLETE

04/14/2023 11:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_