
Midmark M11 - v1.0.0
Total Cycles: 3214

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2023 02:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 183.7 F 3.0
4:00 208.8 F 7.6
6:00 218.2 F 13.3
8:00 228.6 F 8.3
10:00 241.0 F 12.7
12:00 240.6 F 11.4
14:00 238.8 F 9.9
16:00 252.6 F 16.7
18:00 264.3 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.3 F 28.8
2:00 272.3 F 29.1
3:00 271.9 F 28.5
4:00 272.6 F 29.2

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 229.1 F 0.0
10:00 234.9 F 0.0
15:00 231.5 F 0.0
20:00 225.6 F 0.0
25:00 222.0 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

05/10/2023 03:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____