

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001990  
9:21 09/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:58  
PRESSURIZING 2:03  
STERILIZING 3:08  
136.4 C 225kPa 3:08  
136.4 C 226kPa 3:38  
136.4 C 225kPa 4:08  
136.4 C 224kPa 4:38  
136.1 C 223kPa 5:08  
136.2 C 224kPa 5:38  
136.4 C 223kPa 6:08  
136.4 C 226kPa 6:38  
136.4 C 226kPa 6:38  
VENTING 6:38  
AIR DRYING 7:28  
CYCLE COMPLETE 13:53  
-----

Digital Signature #  
4448D1CF7F0C6B8E  
-----