

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3466

Name: RITTER2  
Sterilizer ID: V2313228  
03/21/2024 02:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.9 F 0.0  
2:00 170.0 F 2.2  
4:00 182.1 F 6.2  
6:00 201.6 F 11.0  
8:00 217.4 F 5.2  
10:00 229.5 F 10.1  
12:00 232.8 F 8.6  
14:00 244.1 F 13.4  
16:00 244.1 F 12.7  
18:00 256.8 F 19.4  
20:00 267.8 F 26.5  
22:00 271.7 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.3  
1:00 272.9 F 29.5  
2:00 272.4 F 29.9  
3:00 272.2 F 29.4  
4:00 272.4 F 30.0

Min 271.5 F 29.1  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.9 F 0.0  
5:00 194.8 F 0.0  
10:00 197.8 F 0.0  
15:00 194.8 F 0.0  
20:00 191.4 F 0.0  
25:00 189.0 F 0.0  
30:00 186.6 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 22:03  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:00:58  
CYCLE COMPLETE

03/21/2024 03:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_