

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3285

Name: RITTER 1  
Sterilizer ID: V2178198  
05/25/2023 12:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.5 F 0.0  
2:00 185.7 F 2.5  
4:00 201.2 F 7.6  
6:00 222.5 F 13.5  
8:00 231.5 F 9.0  
10:00 238.3 F 9.5  
12:00 244.5 F 13.0  
14:00 243.8 F 12.1  
16:00 257.4 F 19.3  
18:00 269.2 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.3 F 29.0  
2:00 272.5 F 28.9  
3:00 272.1 F 28.5  
4:00 271.4 F 28.2  
  
Min 271.2 F 28.1  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.9 F 0.0  
5:00 224.4 F 0.0  
10:00 235.6 F 0.0  
15:00 222.8 F 0.0  
20:00 223.1 F 0.0  
25:00 222.7 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 19:07  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:57:56  
CYCLE COMPLETE

05/25/2023 01:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_