

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002235  
14:25 07/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:18  
PRESSURIZING 2:25  
STERILIZING 3:28  
136.0 C 220kPa 3:28  
136.2 C 224kPa 3:58  
136.4 C 225kPa 4:28  
136.2 C 224kPa 4:58  
136.4 C 225kPa 5:28  
136.0 C 221kPa 5:58  
136.4 C 225kPa 6:28  
136.4 C 225kPa 6:58  
136.4 C 225kPa 6:58  
VENTING 6:58  
AIR DRYING 7:49  
CYCLE COMPLETE 37:49

-----  
Digital Signature #  
795361FFF5E40195  
-----