
Midmark M11 - v1.0.5
Total Cycles: 265

Name: RITTER2
Sterilizer ID: V2313228
08/05/2021 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.8 F 0.0
2:00 93.7 F 0.6
4:00 114.1 F 2.5
6:00 143.1 F 5.5
8:00 168.8 F 9.5
10:00 190.2 F 9.1
12:00 212.4 F 6.3
14:00 225.3 F 10.4
16:00 230.7 F 8.0
18:00 239.3 F 11.9
20:00 238.5 F 10.2
22:00 250.8 F 16.3
24:00 261.6 F 22.5
26:00 271.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.4 F 29.4
2:00 272.1 F 29.6
3:00 272.1 F 29.2
4:00 272.4 F 29.8

Min 271.6 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 223.1 F 0.0
10:00 221.5 F 0.0
15:00 213.3 F 0.0
20:00 216.6 F 0.0
25:00 221.6 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 26:34
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:04:44
CYCLE COMPLETE

08/05/2021 09:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____