
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000451
10:54 01/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:20
STERILIZING 4:22
136.0 C 228kPa 4:22
136.4 C 233kPa 4:52
136.0 C 231kPa 5:22
136.4 C 233kPa 5:52
136.3 C 232kPa 6:22
136.3 C 231kPa 6:52
136.1 C 230kPa 7:22
136.3 C 232kPa 7:52
136.3 C 232kPa 7:52
VENTING 7:52
AIR DRYING 8:47
CYCLE COMPLETE 9:38

Digital Signature #
67C2355ABBA36121
