
Midmark M11 - v1.0.5
Total Cycles: 4147

Name: RITTER2
Sterilizer ID: V2313228
11/21/2024 11:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.6 F 0.0
2:00 161.6 F 1.7
4:00 174.2 F 5.0
6:00 200.0 F 8.7
8:00 211.3 F 13.3
10:00 223.4 F 7.2
12:00 229.8 F 11.9
14:00 233.1 F 8.8
16:00 240.9 F 11.6
18:00 244.4 F 13.0
20:00 255.7 F 19.0
22:00 265.9 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.0 F 29.3
2:00 272.2 F 29.6
3:00 271.6 F 29.7
4:00 271.5 F 29.2

Min 271.2 F 29.1
Max 272.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 232.2 F 0.0
10:00 222.3 F 0.0
15:00 222.3 F 0.0
20:00 220.0 F 0.0
25:00 224.8 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:02:53
CYCLE COMPLETE

11/21/2024 12:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____