
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002010
8:43 11/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:37
STERILIZING 3:36
136.0 C 220kPa 3:36
136.3 C 224kPa 4:06
136.4 C 225kPa 4:36
136.4 C 224kPa 5:06
136.4 C 225kPa 5:36
136.4 C 225kPa 6:06
136.2 C 223kPa 6:36
136.3 C 224kPa 7:06
136.3 C 224kPa 7:06
VENTING 7:06
AIR DRYING 8:00
CYCLE COMPLETE 38:00

Digital Signature #
6F3A36DE76AF74CF
