
Midmark M11 - v1.0.5
Total Cycles: 4306

Name: RITTER2
Sterilizer ID: V2313228
01/28/2025 03:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.0 F 0.0
2:00 175.8 F 2.5
4:00 180.1 F 7.0
6:00 202.8 F 12.1
8:00 223.2 F 6.9
10:00 230.5 F 12.2
12:00 234.1 F 9.7
14:00 236.6 F 9.2
16:00 247.8 F 15.0
18:00 260.1 F 21.7
20:00 271.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 271.9 F 29.5
2:00 271.9 F 30.2
3:00 272.1 F 30.0
4:00 272.4 F 30.0

Min 270.9 F 29.2
Max 273.1 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 215.4 F 0.0
10:00 193.8 F 0.0
15:00 191.3 F 0.0
20:00 188.3 F 0.0
25:00 185.0 F 0.0
30:00 182.2 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:00:02
CYCLE COMPLETE

01/28/2025 04:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____