

-----  
Midmark M11 - v1.0.5  
Total Cycles: 415

Name: RITTER2  
Sterilizer ID: V2313228  
10/04/2021 11:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 97.4 F 0.0  
2:00 97.4 F 0.6  
4:00 109.0 F 2.2  
6:00 130.7 F 4.6  
8:00 163.8 F 8.1  
10:00 180.9 F 12.0  
12:00 219.3 F 5.0  
14:00 223.9 F 9.1  
16:00 233.1 F 13.2  
18:00 234.7 F 9.6  
20:00 238.5 F 10.3  
22:00 243.6 F 12.7  
24:00 254.1 F 18.3  
26:00 263.9 F 24.2  
28:00 272.8 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.7  
1:00 272.1 F 30.0  
2:00 272.3 F 30.2  
3:00 272.6 F 30.2  
4:00 272.6 F 30.2

Min 271.7 F 29.6  
Max 273.2 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.3 F 0.0  
5:00 217.5 F 0.0  
10:00 213.2 F 0.0  
15:00 191.9 F 0.0  
20:00 199.8 F 0.0  
25:00 212.9 F 0.0  
30:00 174.8 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 29:01  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:07:18  
CYCLE COMPLETE

10/05/2021 06:16 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01