
Midmark M11 - v1.0.0
Total Cycles: 1912

Name: RITTER 1
Sterilizer ID: V2178198
05/05/2022 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.2 F 0.0
2:00 116.4 F 0.0
4:00 139.8 F 1.1
6:00 164.6 F 3.3
8:00 187.0 F 6.8
10:00 210.1 F 11.0
12:00 216.9 F 4.3
14:00 227.9 F 9.1
16:00 237.9 F 10.9
18:00 240.0 F 11.1
20:00 237.1 F 9.1
22:00 251.3 F 15.9
24:00 263.0 F 22.7
26:00 272.2 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.9
2:00 271.6 F 28.1
3:00 272.7 F 29.1
4:00 272.1 F 28.4

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 211.7 F 0.0
10:00 224.8 F 0.0
15:00 229.8 F 0.0
20:00 230.0 F 0.0
25:00 231.4 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 26:12
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:04:07
CYCLE COMPLETE

05/05/2022 10:04 AM
ACCEPTED: X
REJECTED:
Operator: U02