
Midmark M11 - v1.0.0
Total Cycles: 2363

Name: RITTER 1
Sterilizer ID: V2178198
09/22/2022 03:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 218.0 F 0.0
2:00 211.1 F 4.4
4:00 220.0 F 10.0
6:00 226.6 F 6.1
8:00 235.6 F 11.4
10:00 235.8 F 9.1
12:00 239.0 F 9.4
14:00 248.7 F 14.7
16:00 261.1 F 21.7
18:00 272.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.1
2:00 271.8 F 28.5
3:00 271.6 F 28.7
4:00 272.4 F 29.1

Min 271.2 F 28.2
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 238.8 F 0.0
10:00 232.8 F 0.0
15:00 224.0 F 0.0
20:00 224.2 F 0.0
25:00 229.2 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 2:42
HEATING: 18:30
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

09/22/2022 04:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____