

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002049  
7:45 18/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:46  
STERILIZING 3:52  
136.4 C 221kPa 3:52  
136.3 C 223kPa 4:22  
136.2 C 221kPa 4:52  
136.3 C 223kPa 5:22  
136.1 C 223kPa 5:52  
136.4 C 224kPa 6:22  
136.3 C 222kPa 6:52  
136.2 C 222kPa 7:22  
136.2 C 222kPa 7:22  
VENTING 7:22  
AIR DRYING 8:11  
CYCLE COMPLETE 28:20  
-----

Digital Signature #  
FFA3433C8D374B6A  
-----