
Midmark M11 - v1.0.0
Total Cycles: 4776

Name: RITTER 1
Sterilizer ID: V2178198
07/31/2024 08:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 92.7 F 0.1
4:00 84.6 F 0.2
6:00 125.9 F 2.0
8:00 155.5 F 4.7
10:00 179.9 F 7.9
12:00 205.0 F 11.6
14:00 211.2 F 3.0
16:00 218.5 F 6.6
18:00 228.8 F 10.3
20:00 230.7 F 8.2
22:00 240.3 F 10.8
24:00 234.5 F 8.0
26:00 248.9 F 14.5
28:00 260.0 F 20.7
30:00 270.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.8
2:00 271.5 F 28.2
3:00 272.6 F 28.8
4:00 271.6 F 28.0

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 192.2 F 0.0
10:00 191.8 F 0.0
15:00 195.9 F 0.0
20:00 199.7 F 0.0
25:00 239.2 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 30:53
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:08:57
CYCLE COMPLETE

07/31/2024 09:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____