
Midmark M11 - v1.0.5
Total Cycles: 1848

Name: RITTER2
Sterilizer ID: V2313228
12/15/2022 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.1 F 0.3
2:00 205.1 F 7.1
4:00 222.3 F 11.0
6:00 251.9 F 3.2
8:00 259.5 F 8.6
10:00 276.8 F 13.2
12:00 279.7 F 6.4

C573: HEATUP MODE
STEAM TEMP OVERLIM
CHAMBER TEMP 285F
CHAMBER PRESS 7.4PSI
** ITEMS NOT STERILE **

12/15/2022 01:23 PM