
Midmark M11 - v1.0.5
Total Cycles: 1363

Name: RITTER2
Sterilizer ID: V2313228
08/03/2022 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.5 F 0.0
2:00 127.6 F 0.0
4:00 144.6 F 0.2
6:00 183.6 F 0.8
8:00 202.3 F 4.7
10:00 215.2 F 8.9
12:00 227.9 F 13.5
14:00 232.3 F 9.5
16:00 237.9 F 10.2
18:00 243.3 F 12.7
20:00 242.3 F 11.7
22:00 254.8 F 18.2
24:00 265.8 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.5
2:00 271.7 F 28.7
3:00 272.5 F 29.6
4:00 272.2 F 29.1

Min 271.4 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 212.4 F 0.0
10:00 222.4 F 0.0
15:00 214.4 F 0.0
20:00 212.6 F 0.0
25:00 212.8 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:46
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:03:52
CYCLE COMPLETE

08/03/2022 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____