
Midmark M11 - v1.0.0
Total Cycles: 3613

Name: RITTER 1
Sterilizer ID: V2178198
08/03/2023 02:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.7 F 0.0
2:00 127.0 F 0.0
4:00 150.5 F 1.6
6:00 192.9 F 5.5
8:00 204.9 F 10.6
10:00 217.5 F 4.7
12:00 227.5 F 9.0
14:00 238.4 F 13.0
16:00 238.5 F 10.5
18:00 236.9 F 8.0
20:00 248.8 F 14.7
22:00 259.9 F 20.9
24:00 270.4 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.0
2:00 271.5 F 28.3
3:00 272.7 F 29.2
4:00 271.8 F 28.5

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 211.7 F 0.0
10:00 223.0 F 0.0
15:00 228.3 F 0.0
20:00 231.2 F 0.0
25:00 229.3 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:50
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:58
CYCLE COMPLETE

08/03/2023 03:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____