
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001179
15:48 11/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:20
STERILIZING 4:24
136.4 C 222kPa 4:24
136.3 C 224kPa 4:54
136.4 C 226kPa 5:24
136.3 C 222kPa 5:54
136.1 C 223kPa 6:24
136.3 C 222kPa 6:54
136.4 C 223kPa 7:24
136.4 C 225kPa 7:54
136.4 C 225kPa 7:54
VENTING 7:54
AIR DRYING 8:48
CYCLE COMPLETE 9:06

Digital Signature #
5CDF7C586B2B5650
