
Midmark M11 - v1.0.5
Total Cycles: 1615

Name: RITTER2
Sterilizer ID: V2313228
10/18/2022 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.1 F 0.0
2:00 207.8 F 4.5
4:00 216.9 F 10.8
6:00 225.6 F 7.7
8:00 239.0 F 11.3
10:00 242.4 F 13.4
12:00 244.3 F 13.2
14:00 257.8 F 20.6
16:00 269.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.8 F 30.1
2:00 271.9 F 29.4
3:00 272.6 F 30.3
4:00 272.7 F 30.0

Min 271.6 F 29.3
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 197.3 F 0.0
10:00 216.6 F 0.0
15:00 224.7 F 0.0
20:00 227.3 F 0.0
25:00 221.4 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 3:56
HEATING: 16:58
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:09
CYCLE COMPLETE

10/18/2022 11:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____