
Midmark M11 - v1.0.5
Total Cycles: 3961

Name: RITTER2
Sterilizer ID: V2313228
09/20/2024 08:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.2 F 0.0
2:00 69.8 F 0.0
4:00 73.6 F 0.0
6:00 100.4 F 0.1
8:00 140.9 F 0.6
10:00 168.0 F 2.5
12:00 191.8 F 5.7
14:00 202.8 F 9.5
16:00 218.5 F 13.3
18:00 227.7 F 7.3
20:00 237.2 F 11.5
22:00 233.7 F 8.4
24:00 243.2 F 12.1
26:00 244.4 F 12.7
28:00 256.2 F 18.8
30:00 266.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.1
2:00 271.6 F 29.0
3:00 272.0 F 29.0
4:00 271.7 F 29.1

Min 271.2 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 229.9 F 0.0
10:00 182.3 F 0.0
15:00 176.9 F 0.0
20:00 173.4 F 0.0
25:00 171.8 F 0.0
30:00 170.2 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 31:35
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:10:00
CYCLE COMPLETE

09/20/2024 09:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____