
Midmark M11 - v1.0.0
Total Cycles: 1002

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2021 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.1 F 0.0
2:00 198.8 F 3.9
4:00 199.9 F 8.0
6:00 228.5 F 13.3
8:00 229.6 F 8.3
10:00 240.3 F 12.4
12:00 240.1 F 11.1
14:00 237.9 F 9.4
16:00 251.5 F 16.0
18:00 262.8 F 22.7
20:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 28.9
2:00 271.6 F 28.5
3:00 272.5 F 28.9
4:00 271.5 F 28.3

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 229.6 F 0.0
10:00 228.6 F 0.0
15:00 236.0 F 0.0
20:00 218.8 F 0.0
25:00 221.1 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

07/13/2021 11:26 AM
ACCEPTED: X
REJECTED:
Operator: U01