
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001093
13:55 25/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:36
STERILIZING 3:49
136.0 C 220kPa 3:49
136.2 C 224kPa 4:19
136.4 C 225kPa 4:49
136.4 C 224kPa 5:19
136.1 C 224kPa 5:49
136.4 C 225kPa 6:19
136.3 C 224kPa 6:49
136.4 C 226kPa 7:19
136.4 C 226kPa 7:19
VENTING 7:19
AIR DRYING 8:15
CYCLE COMPLETE 23:15

Digital Signature #
0682E1E986C4ECF3
