
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005650
8:29 02/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:23
PRESSURIZING 1:29
STERILIZING 2:30
136.4 C 220kPa 2:30
136.4 C 223kPa 3:00
136.4 C 220kPa 3:30
136.4 C 223kPa 4:00
136.3 C 219kPa 4:30
136.4 C 220kPa 5:00
136.3 C 221kPa 5:30
136.2 C 220kPa 6:00
136.2 C 220kPa 6:00
VENTING 6:01
AIR DRYING 6:49
CYCLE COMPLETE 8:50

Digital Signature #
7B4CD4AF532B0FBB
