

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3960

Name: RITTER2  
Sterilizer ID: V2313228  
09/19/2024 03:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.6 F 0.0  
2:00 140.4 F 1.4  
4:00 154.3 F 4.4  
6:00 181.6 F 8.5  
8:00 202.5 F 13.0  
10:00 217.5 F 6.4  
12:00 229.9 F 11.4  
14:00 231.2 F 8.3  
16:00 241.0 F 11.3  
18:00 244.6 F 13.2  
20:00 256.8 F 19.7  
22:00 267.9 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.3 F 29.4  
2:00 272.0 F 29.7  
3:00 271.8 F 29.1  
4:00 272.4 F 29.8

Min 271.3 F 29.0  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.9 F 0.0  
5:00 190.0 F 0.0  
10:00 194.3 F 0.0  
15:00 185.6 F 0.0  
20:00 181.0 F 0.0  
25:00 178.2 F 0.0  
30:00 178.6 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 23:18  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:01:47  
CYCLE COMPLETE

09/19/2024 04:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_