

-----  
Midmark M11 - v1.0.5  
Total Cycles: 717

Name: RITTER2  
Sterilizer ID: V2313228  
01/20/2022 03:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
2:00 184.3 F 4.1  
4:00 207.4 F 10.0  
6:00 222.7 F 6.4  
8:00 236.7 F 11.9  
10:00 238.5 F 10.5  
12:00 238.6 F 9.9  
14:00 253.5 F 17.3  
16:00 265.8 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.5 F 29.2  
2:00 271.3 F 28.6  
3:00 272.7 F 29.2  
4:00 271.6 F 28.5  
  
Min 271.2 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
5:00 217.4 F 0.0  
10:00 200.1 F 0.0  
15:00 200.6 F 0.0  
20:00 201.9 F 0.0  
25:00 196.7 F 0.0  
30:00 222.5 F 0.0  
Drying Complete

FILLING: 2:38  
HEATING: 17:37  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:24  
CYCLE COMPLETE

01/20/2022 04:11 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_