
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000797
15:33 16/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 5:40
STERILIZING 7:20
136.4 C 222kPa 7:20
136.4 C 223kPa 7:50
136.1 C 220kPa 8:20
136.1 C 220kPa 8:50
136.4 C 223kPa 9:20
135.9 C 219kPa 9:50
136.4 C 223kPa 10:20
136.4 C 221kPa 10:50
136.4 C 221kPa 10:50
VENTING 10:50
AIR DRYING 11:59
CYCLE COMPLETE 33:17

Digital Signature #
9B2C8EAA09C5BFD1
