
Midmark M11 - v1.0.5
Total Cycles: 2551

Name: RITTER2
Sterilizer ID: V2313228
06/22/2023 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.2 F 0.0
2:00 174.8 F 2.8
4:00 183.2 F 6.8
6:00 205.4 F 11.8
8:00 220.4 F 6.6
10:00 233.2 F 11.8
12:00 234.0 F 9.1
14:00 238.8 F 10.0
16:00 246.9 F 14.1
18:00 258.8 F 20.7
20:00 269.7 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.9
2:00 272.1 F 29.3
3:00 271.3 F 29.0
4:00 272.2 F 29.9

Min 271.3 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 208.0 F 0.0
10:00 218.3 F 0.0
15:00 220.3 F 0.0
20:00 219.4 F 0.0
25:00 218.4 F 0.0
30:00 216.6 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

06/22/2023 10:43 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____