
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000379
14:51 21/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:51
STERILIZING 4:10
136.5 C 232kPa 4:10
136.4 C 233kPa 4:40
136.3 C 233kPa 5:10
136.3 C 233kPa 5:40
136.3 C 233kPa 6:10
136.2 C 231kPa 6:40
136.2 C 231kPa 7:10
136.3 C 233kPa 7:40
136.3 C 233kPa 7:40
VENTING 7:40
AIR DRYING 8:34
CYCLE COMPLETE 11:26

Digital Signature #
0FF8EC53EE795222
