

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1699

Name: RITTER2  
Sterilizer ID: V2313228  
11/08/2022 01:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.8 F 0.0  
2:00 140.6 F 0.2  
4:00 172.2 F 0.7  
6:00 196.9 F 3.8  
8:00 214.3 F 8.6  
10:00 226.9 F 11.2  
12:00 234.5 F 9.8  
14:00 237.3 F 9.2  
16:00 243.6 F 11.6  
18:00 247.9 F 14.1  
20:00 260.9 F 21.3  
22:00 272.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.7 F 28.9  
2:00 271.7 F 28.5  
3:00 272.7 F 28.9  
4:00 271.6 F 28.2

Min 271.5 F 28.2  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.2 F 0.0  
5:00 215.7 F 0.0  
10:00 225.8 F 0.0  
15:00 227.8 F 0.0  
20:00 227.1 F 0.0  
25:00 225.5 F 0.0  
30:00 223.9 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 22:26  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:40  
CYCLE COMPLETE

11/08/2022 02:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_