
Midmark M11 - v1.0.0
Total Cycles: 1618

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2022 01:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 180.5 F 0.0
4:00 207.8 F 0.1
6:00 213.4 F 0.6
8:00 216.8 F 1.5
10:00 222.8 F 3.4
12:00 230.9 F 6.4
14:00 239.5 F 10.0
16:00 240.2 F 10.0
18:00 245.9 F 13.0
20:00 243.1 F 11.7
22:00 240.4 F 10.4
24:00 253.7 F 17.0
26:00 265.9 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.6 F 28.8
2:00 272.8 F 28.7
3:00 272.2 F 28.3
4:00 271.7 F 28.0

Min 271.6 F 27.9
Max 273.0 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 218.2 F 0.0
5:00 235.1 F 0.0
10:00 222.1 F 0.0
15:00 216.2 F 0.0
20:00 209.3 F 0.0
25:00 204.8 F 0.0
30:00 209.6 F 0.0

Drying Complete

FILLING: 1:06
HEATING: 27:45
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:05:57
CYCLE COMPLETE

02/10/2022 03:03 PM
ACCEPTED: X
REJECTED:
Operator: U02