

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 002038  
13:18 13/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:18  
PRESSURIZING 2:22  
STERILIZING 3:25  
136.0 C 222kPa 3:25  
136.0 C 223kPa 3:55  
136.3 C 225kPa 4:25  
136.4 C 226kPa 4:55  
136.2 C 224kPa 5:25  
136.2 C 222kPa 5:55  
136.2 C 225kPa 6:25  
136.0 C 224kPa 6:55  
136.0 C 224kPa 6:55  
VENTING 6:55  
AIR DRYING 7:44  
CYCLE COMPLETE 21:47  
-----

Digital Signature #  
34D965395B52A441  
-----