
Midmark M11 - v1.0.0
Total Cycles: 2339

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2022 02:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 213.4 F 0.0
2:00 208.4 F 4.5
4:00 220.1 F 10.0
6:00 226.3 F 5.9
8:00 235.9 F 11.1
10:00 235.9 F 8.7
12:00 239.9 F 9.7
14:00 248.4 F 14.6
16:00 260.9 F 21.7
18:00 272.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.6 F 29.2
2:00 271.6 F 28.5
3:00 272.3 F 29.3
4:00 272.4 F 29.0

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 230.9 F 0.0
10:00 243.3 F 0.0
15:00 236.8 F 0.0
20:00 226.5 F 0.0
25:00 221.2 F 0.0
30:00 223.0 F 0.0

Drying Complete

FILLING: 2:28
HEATING: 18:30
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

09/19/2022 03:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____