

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3698

Name: RITTER2  
Sterilizer ID: V2313228  
06/13/2024 05:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.5 F 0.0  
2:00 145.7 F 2.4  
4:00 164.0 F 5.6  
6:00 187.0 F 9.8  
8:00 210.9 F 6.6  
10:00 223.3 F 8.3  
12:00 236.4 F 13.2  
14:00 237.1 F 10.3  
16:00 235.6 F 8.7  
18:00 248.7 F 15.1  
20:00 260.2 F 21.6  
22:00 270.9 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.6 F 29.6  
1:00 272.9 F 29.7  
2:00 273.0 F 29.8  
3:00 273.1 F 30.0  
4:00 272.7 F 30.1

Min 271.7 F 29.5  
Max 273.3 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.8 F 0.0  
5:00 236.6 F 0.0  
10:00 218.6 F 0.0  
15:00 182.5 F 0.0  
20:00 178.9 F 0.0  
25:00 176.0 F 0.0  
30:00 175.4 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 23:16  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:02:04  
CYCLE COMPLETE

06/13/2024 06:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_