
Midmark M11 - v1.0.5
Total Cycles: 1336

Name: RITTER2
Sterilizer ID: V2313228
07/25/2022 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.6 F 0.0
2:00 177.6 F 0.1
4:00 190.6 F 2.7
6:00 208.0 F 7.6
8:00 232.1 F 13.0
10:00 232.2 F 9.7
12:00 235.7 F 9.0
14:00 243.7 F 12.5
16:00 244.5 F 12.9
18:00 256.9 F 19.6
20:00 267.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.4 F 29.8
2:00 271.6 F 29.0
3:00 272.1 F 29.9
4:00 272.4 F 29.5

Min 271.1 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 221.8 F 0.0
10:00 235.0 F 0.0
15:00 219.9 F 0.0
20:00 216.9 F 0.0
25:00 222.9 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

07/25/2022 10:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____