

-----  
Midmark M11 - v1.0.5  
Total Cycles: 208

Name: RITTER2  
Sterilizer ID: V2313228  
07/14/2021 12:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.4 F 0.0  
2:00 161.7 F 2.8  
4:00 166.2 F 6.5  
6:00 191.8 F 11.2  
8:00 207.3 F 5.1  
10:00 222.2 F 10.3  
12:00 229.5 F 8.4  
14:00 240.7 F 13.5  
16:00 242.4 F 12.5  
18:00 254.7 F 18.9  
20:00 265.7 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.5  
1:00 272.4 F 30.2  
2:00 272.0 F 29.6  
3:00 271.4 F 29.7  
4:00 272.2 F 30.3

Min 271.2 F 29.3  
Max 272.6 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 220.6 F 0.0  
10:00 227.6 F 0.0  
15:00 228.4 F 0.0  
20:00 219.1 F 0.0  
25:00 216.3 F 0.0  
30:00 187.4 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 21:49  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:32  
CYCLE COMPLETE

07/14/2021 01:32 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01