

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 006899  
14:06 02/02/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:44  
PRESSURIZING 2:47  
STERILIZING 3:52  
136.4 C 220kPa 3:52  
136.4 C 225kPa 4:22  
136.4 C 222kPa 4:52  
136.4 C 224kPa 5:22  
136.4 C 222kPa 5:52  
136.2 C 221kPa 6:22  
136.5 C 225kPa 6:52  
136.4 C 223kPa 7:22  
136.4 C 223kPa 7:22  
VENTING 7:22  
AIR DRYING 8:16  
CYCLE COMPLETE 26:56

-----  
Digital Signature #  
DD3616F3C82DF597  
-----