

-----  
Midmark M11 - v1.0.5  
Total Cycles: 902

Name: RITTER2  
Sterilizer ID: V2313228  
03/24/2022 11:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.2 F 0.0  
2:00 184.8 F 2.9  
4:00 208.3 F 8.7  
6:00 229.7 F 9.1  
8:00 239.5 F 11.7  
10:00 238.5 F 10.4  
12:00 238.2 F 10.0  
14:00 253.6 F 17.6  
16:00 266.5 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.8 F 29.6  
2:00 272.6 F 29.5  
3:00 272.7 F 29.5  
4:00 272.4 F 29.6  
  
Min 270.8 F 28.6  
Max 273.1 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
5:00 168.6 F 0.0  
10:00 168.3 F 0.0  
15:00 163.4 F 0.0  
20:00 163.4 F 0.0  
25:00 162.7 F 0.0  
30:00 164.3 F 0.0  
Drying Complete

FILLING: 2:07  
HEATING: 17:20  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:56:38  
CYCLE COMPLETE

03/24/2022 01:25 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02