
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000050
10:35 14/09/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 6:10
PRESSURIZING 8:12
STERILIZING 10:12
136.0 C 223kPa 10:12
136.1 C 224kPa 10:42
136.4 C 227kPa 11:12
136.4 C 226kPa 11:42
136.4 C 226kPa 12:12
136.4 C 228kPa 12:42
136.3 C 227kPa 13:12
136.4 C 226kPa 13:42
VENTING 13:42
AIR DRYING 14:56
CYCLE COMPLETE 44:56

Digital Signature #
957C17C30462CD76
