
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001690
16:42 24/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:47
STERILIZING 3:52
136.4 C 224kPa 3:52
136.4 C 225kPa 4:22
136.4 C 226kPa 4:52
136.3 C 223kPa 5:22
136.4 C 224kPa 5:52
136.4 C 226kPa 6:22
136.4 C 224kPa 6:52
136.4 C 225kPa 7:22
136.4 C 225kPa 7:22
VENTING 7:22
AIR DRYING 8:15
CYCLE COMPLETE 38:15

Digital Signature #
2C36C1671DC84F10
