

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1668

Name: RITTER 1  
Sterilizer ID: V2178198  
02/24/2022 11:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.7 F 0.0  
2:00 190.8 F 2.5  
4:00 211.2 F 7.5  
6:00 224.2 F 12.2  
8:00 231.8 F 8.6  
10:00 239.9 F 10.9  
12:00 243.4 F 11.8  
14:00 241.1 F 10.5  
16:00 255.2 F 17.6  
18:00 267.1 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 27.8  
1:00 272.6 F 28.6  
2:00 271.5 F 27.7  
3:00 272.6 F 28.6  
4:00 271.7 F 27.8  
  
Min 271.2 F 27.7  
Max 272.7 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.5 F 0.0  
5:00 236.6 F 0.0  
10:00 227.5 F 0.0  
15:00 220.6 F 0.0  
20:00 179.8 F 0.0  
25:00 172.6 F 0.0  
30:00 171.2 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 19:16  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 00:58:21  
CYCLE COMPLETE

02/24/2022 12:29 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_